

<b>Product Code:</b>	A9742
<b>Product Name:</b>	PCR Cyclers Validation Kit
<b>Short Description:</b>	Usage: for conventional PCR block cyclers
<b>Specifications:</b>	Annealing temp. 63°C: two amplicons Annealing temp. 65°C: one amplicon Annealing temp. 68°C: no amplicon
<b>WGK:</b>	nwg
<b>Storage:</b>	2-8°C
<b>Shipment:</b>	wet ice
<b>CS:</b>	38220000
	